

**GOVERNMENT OF ASSAM**  
**OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM**  
**SIXMILE, KHANAPARA, GUWAHATI-22**  
*Tele-Fax No.0361-2366236, E-mail- [dmeassam@gmail.com](mailto:dmeassam@gmail.com), [dme@assam.gov.in](mailto:dme@assam.gov.in),*

---

No.DME/284/2014/11370,

Dated : 14 -08-2015.

**CLARIFICATION NOTICE**

**IFB No. No. DME/284/2014/10142.**

**Dated 21-07-15**

1. **Longevity Projection** : Longevity projection of different manufacturers of pacemakers are calculated from different sets of values. Hence we have given a range of values rather than fixed parameters for calculation of Longevity Projection with default parameter on. Longevity Projection has limited value in pacing system selection, as these values are derived in vitro only with no consideration of patient related factors.
2. **Auto-Capture**: By the word Auto-Capture we did not mean Trade-mark term of any manufacturer. Any pacemaker system which measures pacing threshold and adjusts pacing output accordingly will qualify for this term. However considerations will be given on the nature of output adjustment and their impact on pacemaker function of different products during technical evaluation. We have not made it a mandatory feature.
3. **Warranty**: As regards to the warranty terms, the bid participants are to quote their own Warranty provisions. How they will be evaluated will be decided by the Technical Committee when it meets after opening of the technical bids. Evaluation of any of the quoted parameters can not be a subject of open discussion at this stage.

Other terms & conditions remain same.

Sd/- (Prof. (Dr.) B. P. Chakravarty)  
Director of Medical Education, Assam