

MINUTES OF PRE-BID MEETING HELD ON 27-07-2020 AT 11:00 AM IN ROOM NO 4C (2ND FLOOR), O/O DIRECTORATE OF MEDICAL EDUCATION, ASSAM REGARDING THE TENDER FOR SUPPLY AND INSTALLATION OF VARIOUS EQUIPMENT TO LAKHIMPUR MEDICAL COLLEGE & HOSPITAL, NORTH LAKHIMPUR IN RESPECT OF TENDER VIDE IFB NO. DME/144/2020/7225, DATED 26-06-2020.

Members Present:

1. Dr. S. R. Barman
Principal-cum-Chief Superintendent, DMC&H, Diphu
2. Dr. H. Choudhury
Assistant Professor , Radiology , GMCH, Guwahati
3. Dr. Pran Pratim Saikia
Demonstrator, Pharmacology., GMCH, Guwahati

Documents placed before Committee:

1. Tender document.
2. Complaints/suggestions received through e-mail/hand copy etc. from prospective bidders.

Evaluation:


- A. Complaint raised/suggestions filed by bidders/prospective bidders:

Sl. No.	Name of prospective bidder/firm	Query/complaint/suggestion (As per Annexure)	Decision taken by Committee with reasoning
1.	M/s United Traders	Annexure-1	Sr. no. 46 (Suction Machine Electric) is dropped from the current tender as it has already approved rate list
2.	M/s ItasG Pharma	Annexure-2	Decision made as follows

Notes:

1. Sr. no. 121,122,126, 128, 136, 208, 213, are repeated hence, dropped from the tender list and quantity required shall to be adjusted.
2. Name of Sl no. 61 is corrected as "O2 oxygen cylinder carrier".
3. Name of Sl. no 169. is corrected as "Snellen chart/ snellen drum".
4. Name of Sl no. 177 is corrected as "Otoscope with halogen bulb & siegal attachment".
5. Name of Sl no. 221 is corrected as Microtome with knife".
6. Name of Sl no. 234 is corrected as "Dissection set complete".
7. Name of Sl no. 240 is corrected as "Infrared spectroscope".
8. Name of Sl no. 249 is corrected as "Tooth Extractor left and right".
9. Name of Sl no. 296 is corrected as "Brain knife".
10. Sl. No. 101,174,175,176,95,96,178,79,180,182,83,184,33,185,186,60 are dropped from the list and quantity required shall to be adjusted.


27/07/2020
(Dr. S. R. Barman)


27.07.20
(Dr. H. Choudhury)


27/7/20
(Dr. Pran Pratim Saikia)

UNITED TRADERS

R.K. Bordoloi Road,
Dibrugarh - 786001
(ASSAM)

Mobile No: 9435330039
Email ID: unitedtraders21@gmail.com

Date: 15/07/2020

To,
The Director of Medical Education, Assam,
Sixmile, Khanapara,
Guwahati-22

Sub: Request for dropping of Item No: 46 – Suction Machine Electric

Ref: Approved List No: DME/AppvList/312/2019/6433-46 Dt: 08/06/2020

Ref: IFB NO.DME/144/2020/7225 Dated: 26/06 /2020 TENDER FOR SUPPLY & INSTALLATION OF VARIOUS EQUIPMENT / INSTRUMENT FOR LAKHIMPUR MEDICAL COLLEGE & HOSPITAL, LAKHIMPUR ,ASSAM UNDER THE HEALTH & E.W.(B) DEPARTMENT, GOVT. OF ASSAM

Sir,

With reference to the above Tender Notice, we would like to state to you that **Item No: 46 - Suction Machine Electric** has already been approved in the earlier tender vide approved list no: **Approved List No: DME/AppvList/312/2019/6433-46 Dt: 08/06/2020**. (Approved rates for the supply & installations of various equipments related to Department of Biochemistry & ICU).

Here we would like to mention that Suction Machine Electric as demanded in the tender and Electrical Suction Machine for which we are under rate contract is the same product, as per approved list details mentioned above and hence we request you to kindly drop the said Item from the above Tender as the item is already approved in our favour.

Thanking You,
Yours faithfully,
For United Traders,

Proprietor





ITAS G PHARMA

House 7, 3rd Floor, Masjid Galt, Sree Nagar, Guwahati, Assam 781005, India
operations@itaspgharma.in | service@itaspgharma.in
+91 36123 44692
D.L.# D/DL/KMP/19276 | D/DL/KMP/19277
18ACKPD6844P22X
1404000330

Ref.: IGP/2020-2021/288

Date: -21.07.2020

To,
The Director of Medical Education, Assam,
Six Mile, Khanapara, Guwahati-22.

Sub: - Request for amendments / clarifications of the Tender.

Ref: - (i) IFB No. DME/145/2020/7220; Dated: 26.06.2020
(ii) IFB No. DME/145/2020/7225; Dated: 26.06.2020

Respected Sir,

With reference to the above referred tender enquiries, we like to bring to your notice that there are items which are common in both the tenders. The details of such items are listed below: -

1. List of the repeated items: -

Sl. No.	Item Name	Sl. No. as per Sch-03 of IFB No.7220; Dated 26.06.2020	Sl. No. as per Sch-03 of IFB No.7225; Dated 26.06.2020
01.	Laryngeal mirrors different sizes	8	101
02.	Ear suction different sizes	11	174
03.	Nasal suction different sizes	12	175
04.	Siegel's speculum	14	176
05.	Tuning fork (512 Hz)	15	95
06.	Tuning fork (256 Hz)	16	96
07.	Tuning fork (1024 Hz)	17	97
08.	Bayonet forceps	19	178
09.	Haemostatic forceps	20	179
10.	Tilles nasal packing forcep	21	180
11.	Foreign body hook	25	182
12.	Tympanometry set	26	183
13.	Aural Syringe	27	184
14.	Kidney Tray	28	33, 34, 35, 157
15.	Blopsy forcep- (punch & cup)	29	185
16.	Posterior rhinoscopic mirror	30	186
17.	Instrument Tray	31	60,61,154, 155, 156

Ranjit



ITAS G PHARMA

House 7, 3rd Floor, Masjid Gali, Sree Nagar, Guwahati, Assam 781005, India

operations@itaspgharma.in | service@itaspgharma.in

+91 36123 44692

D.L.# D/OL/KMP/19276 | D/OL/KMP/19277

18ACKP06844P27X

1404000330

Further, it has been observed that specifications of some of the tendered items are not present in respect to the minimum requirements of the end user. Minimum specifications to be mentioned for attaining proper comparison and selection of acceptable product range. The details of such items are listed below. We are however, providing hereunder the technical specifications of some items which have gathered.

2. List of requirements of Specifications/size of the following items: -

Sl No.	Item Name	Sl. No. as per Sch-03 of IFB No.7220; Dated 26.06.2020	Sl. No. as per Sch-03 of IFB No.7225; Dated 26.06.2020	Specifications offered.
1	Sponge Holder Forcep	N/A	16	Sponge Holder Forcep 25cm
2	Ovum Forcep	N/A	17	Ovum Forcep 25cm
3	Cheattle's Forcep	N/A	18	Cheattle's Forcep 25cm
4	Epiotomy Forcep	N/A	20	Epiotomy Forcep 14cm
5	Sim's Speculum	N/A	29	Sim's Speculum SMALL
6	Adsons Dissecting Forcep	N/A	114	Adsons Dissecting Forcep 12.5cm
7	Mosquito Artery Forcep straight	N/A	121	Mosquito Artery Forcep straight 12.5cm
8	Bayonet Forceps	19	178	Bayonet Forceps 18cm
9	Haemostatic Forceps	20	179	Haemostatic Forceps 12.5cm
10	Tilles Nasal Packing Forcep	21	180	Tilles Nasal Packing Forcep 14cm
11	Cautery Machine	N/A	191	Input Voltage: 180-270 VAC Current Fuse - 7 Amps Analog control of Power output with LED Indicator of different modes and push switch control-ANALOG 250 W. The unit has to be IEC 60601-2-2 & 60601-1. Protection Class -2B, Unit Type CF, Nominal frequency : Different for different modes. High Frequency : < 150 μ A r.m.s. in all output mode. Tone indications to differentiate the mode of applications and adjustable in 03 stages. Weight- 4.5 kgs. Size: 295 x 120 x 285 mm (L X H X W)
12	Toothed Vulsillum	N/A	210	Toothed Vulsillum 24cm
13	Crocodile Forceps	N/A	N/A	Crocodile Forceps 8cm
14	Biopsy Forcep- (Punch & Cup)	12	N/A	Biopsy Forcep- (Punch & Cup) 24.5 cm
15	Otoendoscope	32	N/A	Specifications required.
16	Nasal endoscope	33	N/A	Specifications required.

Rangjit



ITAS G PHARMA

House 7, 3rd Floor, Masjid Gali, Sree Nagar, Guwahati, Assam 781005, India

operations@itasgpharma.in | service@itasgpharma.in

+91 36123 44692

D.L.# D/OL/KMP/19276 | D/OL/KMP/19277

18ACKPD6844P2ZX

1404009330

Sl No.	Item Name	Sl. No. as per Sch-03 of IFB No.7220; Dated 26.06.2020	Sl. No. as per Sch-03 of IFB No.7225; Dated 26.06.2020	Specifications offered.
17	Laryngeal Telescope	34	N/A	Specifications required.
18	Mobile Spot Light	36	N/A	Adjustable on telescopic SS Pipe ranging from 4 ft to 6 ft. Light intensity at 0.5 m : 60,000 Lux. No. of LED bulbs: 07 Nos. Dome Dia 9". Colour Temperature 4800°K. Light Field Diameter 100 mm. Control on dome : On/Off.
19	Instrument Tray with Cover.	31	N/A	Stainless Steel 304 grade Size: 12" x 10", 15" x 12", 9" x 6"
20	Kidney Tray:	28	157	Stainless Steel 304 grade. Size: 12", 10", 8"

So, we request your goodself to issue the necessary amendments / clarifications in view of the above for wider participation and justification of the entire tender process.

Thanking you,

Yours faithfully,

FOR ITAS G. PHARMA


Manager